



# ENV AUTOMATION PTE LTD

60 Paya Lebar Road, #07-54 Paya Lebar Square, Singapore 409051  
Email: envautomation.sg@gmail.com | Website: www.autosiltstopper.com

CERTIFICATE NO  
**ENV2026030**

## CALIBRATION CERTIFICATE

This certificate confirms that the instrument listed below has been inspected and calibrated according to internal inspection requirements and applicable calibration practice.

### CLIENT INFORMATION

KTC Civil Engineering & Construction Pte Ltd

### PROJECT NAME

CR206-ECM2

### SENSOR TYPE

TSS Sensor

### SENSOR NUMBER

25280570030S

### DATE OF CALIBRATION

12/03/2026

### NEXT CALIBRATION

12/03/2027 VALID

Test Items	Test Requirements	Test Result
Visual inspection test	No obvious scratches, damage, ect	<span>Qualified</span>
Does the screw have a leak lock	There is no	<span>Qualified</span>
There is no sundry in the sensor	There is no	<span>Qualified</span>
There is no sundry in the sensor	Soak for more than 12 hours without water	<span>Qualified</span>
There is no sundry in the sensor	Connect the transmitter OK	<span>Qualified</span>
There is no sundry in the sensor	High temperature test for 3 hours the above	<span>Qualified</span>
There is no sundry in the sensor	Oscillate 250 rpm for more than 2 hours	<span>Qualified</span>

Functional Test	Standard Solution		Actual Measurement	
Calibration test (mg / L : 10 % 10 mg /L) (Standard)	Solution 1	0.2	0.53	0.33
	Solution 2	207	204.97	1.0 %
	Solution 3	5070	5012.49	1.1 %
	Solution 4	12050	12237.36	1.6 %

Note : Before testing must refer to the customer requirement and inspection standards.

### AUTHORIZED BY

Checked by Signature

### OFFICIAL COMPANY STAMP

